

**OnSite® COVID-19 Ag Rapid Test**

REF R0182C CE

**Instructions for Use**



Barcode for RTR Use Only

**INTENDED USE**

The OnSite COVID-19 Ag Rapid Test is a lateral flow immunoassay for the qualitative detection of SARS-CoV-2 nucleocapsid antigens in nasopharyngeal (NP) or nasal swab specimens from individuals suspected of COVID-19, within the first seven days of the onset of symptoms. The test is intended for use by healthcare providers or personnel trained in rapid test procedure, as an aid in identifying SARS-CoV-2 infection.

The OnSite COVID-19 Ag Rapid Test does not differentiate between SARS-CoV and SARS-CoV-2.

Positive results indicate the presence of viral antigens, but clinical correlation with patient history and other diagnostic information is necessary to determine infection status. Positive results do not rule out other bacterial or viral infections.

Negative results from patients with symptom onset beyond seven days should be confirmed with a molecular assay. Negative results do not rule out SARS-CoV-2 infection and should not be used as the sole basis for treatment or patient management decisions. Negative results should be considered in the context of a patient's recent exposures, history and the presence of clinical signs and symptoms consistent with COVID-19.

The product is intended to be used in any laboratory and non-laboratory environment that meets the requirements specified in the Instructions for Use and local regulations. For *in vitro* diagnostic use only.

**SUMMARY AND EXPLANATION OF THE TEST**

SARS-CoV-2 belongs to the broad family of coronaviruses which are capable of causing illnesses ranging from the common cold to more severe diseases<sup>1</sup>. SARS-CoV-2 infections cause COVID-19 disease resulting in a wide range of clinical symptoms, ranging from asymptomatic to fever, tiredness and dry cough, and possibly leading to severe sickness and even death. Most patients recover without special treatment. According to recent data, approximately 15-20% of infected individuals become seriously ill and develop difficulty breathing<sup>2</sup>. The elderly and those with underlying medical problems, such as high blood pressure, heart problems or diabetes are more likely to develop serious illness<sup>2</sup>.

Human-to-human transmission of the virus has been confirmed and occurs primarily via respiratory droplets from coughs and sneezes within a range of about six feet (1.8 m)<sup>3</sup>. Viral RNA has also been found in stool samples from patients. It is possible that the virus can be infectious even during the incubation period, but this has not yet been proven<sup>4</sup>.

The current laboratory method for detecting COVID-19 is PCR. However, this method requires sophisticated equipment and highly trained laboratory technicians. The OnSite COVID-19 Ag Rapid Test is an easy-to-use and cost-efficient assay that can be performed at point-of-care settings.

The OnSite COVID-19 Ag Rapid Test detects the presence of antigens from the SARS-CoV-2 virus within the first seven days of the onset of symptoms. Test results should be interpreted at 15 minutes. Results should not be interpreted after 20 minutes. Minimally skilled personnel can perform the test, without the use of cumbersome laboratory equipment.

**TEST PRINCIPLE**

The OnSite COVID-19 Ag Rapid Test is a lateral flow chromatographic immunoassay. The test cassette consists of: 1) a colored conjugate pad containing anti-SARS-CoV-2 antibodies conjugated with colloidal gold (antibody conjugates) and 2) a nitrocellulose membrane strip containing a test line (Ag line) and a control line (C line). The test line is pre-coated with anti-SARS-CoV-2 antibodies and the C line is pre-coated with control antibodies.

The specimen is collected with a nasopharyngeal or nasal swab and the SARS-CoV-2 antigen is extracted from the swab with extraction buffer. Alternatively, samples stored in viral transport medium (VTM) can be directly tested. When applied to the sample well, the extracted specimen migrates across the test strip by capillary action. SARS-CoV-2 antigen, if present in the extract, binds to the antibody conjugates and the immunocomplex is then captured on the membrane by the pre-coated anti-SARS-CoV-2 antibody, forming a colored Ag line that indicates a COVID-19 positive test result.

The test contains an internal control (C line), which should exhibit a colored line regardless of color development on the Ag line. If the C line does not develop, the test result is invalid and the specimen must be retested with a new device.

**REAGENTS AND MATERIALS PROVIDED**

1. Individually sealed foil pouches containing:
  - a. One cassette device
  - b. One desiccant
2. Sealed pouch containing pre-filled extraction tubes
3. Extraction tube nozzles
4. Extraction tube rack
5. Individually sealed pouches containing a sterile swab
6. Instructions for Use

**MATERIALS REQUIRED BUT NOT PROVIDED**

1. Clock, watch or other timing device
2. Disposable gloves, biohazard disposal container

**WARNINGS AND PRECAUTIONS**

**For In Vitro Diagnostic Use**


1. Read these Instructions for Use completely before performing the test. Failure to follow these instructions could lead to inaccurate test results.
2. Do not open the sealed pouch unless ready to conduct the assay.
3. Do not use expired devices.
4. Bring all reagents to room temperature (15-30°C) before use.
5. Do not use the components in any other type of test kit as a substitute for the components in this kit.
6. Wash hands thoroughly before and after testing. We recommend wearing disposable gloves while handling kit reagents and clinical specimens.
7. Do not smoke, drink or eat in areas where specimens or kit reagents are being handled.
8. Dispose of all specimens and materials used to perform the test as biohazardous waste.
9. Read the test results 15 minutes after specimen is applied to the sample well. Consider any results read after 20 minutes invalid and repeat the test.
10. Do not perform the test in a room with strong air flow, i.e. an electric fan or strong air-conditioning.

**REAGENT PREPARATION AND STORAGE INSTRUCTIONS**

All reagents are ready to use as supplied. Store unused devices unopened at 2-30°C. If stored at 2-8°C, ensure that the device is brought to room temperature before opening. The cassette device is stable until the expiration date printed on the sealed pouch. Do not freeze the kit or expose the kit to temperatures above 30°C.

**PREPARATION FOR ASSAY PROCEDURE**

1. Before running the assay, ensure the test area is sanitized. Open the kit and ensure all materials described in "Reagents and Materials Provided" are included and the kit is not expired. Obtain a timing device (clock, watch or timer) and read the Quick Reference Guide and these Instructions for Use.
2. Wash or sanitize hands thoroughly.
3. Fold/assemble the sample extraction tube rack.
4. Remove one pre-filled extraction tube from the sealed pouch and close the pouch with the unused tubes. Hold the pre-filled extraction tube upright and, before opening it, tap the bottom of the tube on a clean, flat surface to ensure that any liquid on the seal is moved down into the tube.
5. Carefully remove the foil seal from the extraction tube, and place the open tube in the sample extraction tube rack provided with the kit.




**SAMPLE COLLECTION**

Consider any materials of human origin as potentially infectious, and handle them with standard biosafety procedures.

1. Remove mucus from the patient's nose.
- Nasal swab specimens**
- 2a. Hold the patient's head in a vertical position and looking slightly downwards (see Figure A).
- 3a. Open the swab package.
 

**Note: Do not touch the swab's absorbent tip, so be sure to open the package on the opposite end.**
- 4a. Carefully insert the entire absorbent tip of the swab in one nostril and rotate at least 5 times. **Be sure that the absorbent tip of the swab scrapes against the nasal wall. Stop when you feel resistance (no more than 2-3 cm into the nose).**
- 5a. Remove swab from nostril and, using the same swab, repeat step 4a in the other nostril.
- 6a. Withdraw the swab from the nasal cavity. Insert the absorbent tip of the swab into the extraction buffer tube and swirl the swab at least 5 times. Proceed to specimen extraction following the Assay Procedure described below.




**Nasopharyngeal swab specimens**

- 2b. Hold the patient's head in a vertical position and looking slightly upwards (see Figure B).
- 3b. Open the swab package.
 

**Note: Do not touch the swab's absorbent tip, so be sure to open the package on the opposite end.**
- 4b. Carefully insert the entire absorbent swab tip into the patient's nostril that presents the most secretion, keeping it near the nose septum floor while gently pushing into the posterior nasopharynx.
- 5b. Rotate the swab at least 5 times.
- 6b. Withdraw the swab from the nasal cavity. Insert the absorbent tip of the swab into the extraction buffer tube and swirl the swab at least 5 times. Proceed to specimen extraction following the Assay Procedure described below.

**ASSAY PROCEDURE**

1. Squeeze the tube against the submerged swab several times to facilitate extraction of the specimen. Remove the swab, place it back in its original wrapping and dispose into a biohazard disposal container.
2. Place the nozzle onto the extraction tube and ensure it is attached firmly.
3. Remove the cassette device from the sealed pouch just prior to testing. Lay the device on a clean, flat surface and label with specimen ID/name.
4. Invert the sample extraction tube and **slowly add 3 drops of the extracted specimen into the sample well of the cassette device by gently squeezing the sample tube.**
5. Set the timing device for 15 minutes.
6. Read the results after 15 minutes.



**Note: The result might be visible after a shorter time, however, it should only be interpreted between 15-20 minutes after dispensing the sample material onto the cassette device.**

Collect all used items (swab, cassette, sample extraction tube, foil seal and nozzle, and used gloves) and discard as biohazardous waste following local laws governing the disposal of devices.

**INTERPRETATION OF ASSAY RESULT**

**NEGATIVE RESULT:** If only the C line develops, the test did not detect SARS-CoV-2 virus (antigen) is present in the specimen. The result is negative or non-reactive.

**POSITIVE RESULT:** If both the C line and Ag line develop, SARS-CoV or SARS-CoV-2 virus (antigen) is detected in the specimen. The result is positive or reactive. Some specimens might produce a faint band, but every visible test line band indicates a positive result independent of the band intensity.

**INVALID:** If no C line develops, the assay is invalid regardless of color development on the Ag line. Repeat the assay with a new device.



**QUALITY CONTROL**

- Internal Control:** This test contains a built-in control feature, the C line. If the C line does not develop after sample application, the result is invalid. Review the entire procedure and repeat the test with a new device.
- External Control:** Good Laboratory Practice recommends using external controls, positive and negative, to ensure the proper performance of the assay, particularly under the following circumstances:
  - A new operator uses the kit, prior to performing the testing of specimens.
  - A new lot of test kits is used
  - A new shipment of test kits is used
  - The temperature during storage of the kits falls outside of 2-30°C.
  - The temperature of the test area falls outside of 15-30°C.
  - To verify a higher than expected frequency of positive or negative results.
  - To investigate the cause of repeated invalid results.

**PERFORMANCE CHARACTERISTICS**

**1. Clinical Performance**

**1.1. Clinical performance in nasopharyngeal swab specimens**

The clinical performance of the OnSite COVID-19 Ag Rapid Test was evaluated at three clinical sites (Colombia, China, and India) in nasopharyngeal (NP) swabs specimens collected from subjects suspected of COVID-19 and from healthy individuals. Two NP swabs were collected from each subject, one for testing by the OnSite COVID-19 Ag Rapid Test and the other one for testing by commercially available real-time Polymerase Chain Reaction (RT-PCR) assay for the detection of SARS-CoV-2, used as the reference method for this study. The performance of the OnSite COVID-19 Ag Rapid Test in this study is shown on the table below:

RT-PCR Test (Reference)	OnSite COVID-19 Ag Rapid Test Result		
	Positive	Negative	Total
Positive	60	5	65
Negative	0	370	370
Total	60	375	435

Relative Sensitivity: 92.3% (95% CI: 83.0-97.5%); Relative Specificity: 100% (95% CI: 99.0-100%); Overall Agreement: 98.9% (95% CI: 97.3-99.6%)

**1.2. Clinical performance in nasal swab specimens**

The clinical performance of the OnSite COVID-19 Ag Rapid Test was evaluated at three clinical sites (Colombia, Brazil, and India) in nasal swabs specimens collected from subjects suspected of COVID-19 and from healthy individuals. Two nasal swabs were collected from each subject, one for testing by the OnSite COVID-19 Ag Rapid Test and the other one for testing by commercially available real-time Polymerase Chain Reaction (RT-PCR) assay for the detection of SARS-CoV-2, used as the reference method for this study. The performance of the OnSite COVID-19 Ag Rapid Test in this study is shown on the table below:

RT-PCR Test (Reference)	OnSite COVID-19 Ag Rapid Test Result		
	Positive	Negative	Total
Positive	36	2	38
Negative	0	170	170
Total	36	172	208

Relative Sensitivity: 94.7% (95% CI: 82.3-99.4%); Relative Specificity: 100% (95% CI: 97.0-100%); Overall Agreement: 99.0% (95% CI: 96.6-99.9%)

**2. Analytical Performance**

**2.1 Analytical Sensitivity (Limit of Detection, LoD)**

The LoD of the OnSite COVID-19 Ag Rapid Test was determined by evaluating a serial dilution of Gamma-Irradiated SARS-CoV-2 virus lysate (BEI Resources, NR-52287). Multiple negative nasopharyngeal or nasal swab specimens were eluted in PBS and were combined and mixed thoroughly to create clinical negative matrix pools for each matrix, to be used as the diluent. Inactivated SARS-CoV-2 virus lysate was diluted in each of these matrices to generate virus dilutions for testing. Each NP or nasal swab was spiked with 50 µL of each virus dilution, extracted with extraction buffer and tested according to the product IFU. The assay LoD was determined for both NP and nasal swab specimens as the lowest concentration that was detected ≥ 95% of the time in the respective specimen matrix.

The LoD of the OnSite COVID-19 Ag Rapid Test in both nasopharyngeal and nasal swab matrices was determined to be 280 TCID<sub>50</sub>/mL. The OnSite COVID-19 Ag Rapid Test can detect the U.K. South Africa and Brazil variants at similar levels as the original SARS-CoV-2 strain.

**2.2 Analytical Specificity (Cross-Reactivity and Microbial Interference)**

The analytical specificity of the OnSite COVID-19 Ag Rapid Test was evaluated by testing commensal and pathogenic microorganisms that may be present in the nasal cavity. Each of the organisms was tested in triplicate in the presence of 2-3X LoD recombinant SARS-CoV-2 NP antigen. No cross-reactivity (except SARS-coronavirus) or interference were seen with the following microorganisms when tested at the concentration presented in the table below:

Potential Cross-Reactant	Concentration	Cross-Reactivity (Yes/No)	Microbial Interference (Yes/No)
SARS-coronavirus NP antigen	25 µg/mL	Yes (3/3 positive)	No (3/3 positive)
MERS-coronavirus NP antigen	25 µg/mL	No (3/3 negative)	No (3/3 positive)
Human coronavirus HKU1 NP antigen	66 µg/mL	No (3/3 negative)	No (3/3 positive)
Human coronavirus 229E	1.77×10 <sup>5</sup> TCID <sub>50</sub> /mL	No (3/3 negative)	No (3/3 positive)
Human coronavirus OC43	0.53×10 <sup>5</sup> TCID <sub>50</sub> /mL	No (3/3 negative)	No (3/3 positive)
Human coronavirus NL63	0.51×10 <sup>5</sup> TCID <sub>50</sub> /mL	No (3/3 negative)	No (3/3 positive)
Adenovirus	7×10 <sup>8</sup> NIU/mL	No (3/3 negative)	No (3/3 positive)
Human Metapneumovirus (hMPV)	0.76×10 <sup>4</sup> TCID <sub>50</sub> /mL	No (3/3 negative)	No (3/3 positive)
Parainfluenza virus 1	5.01×10 <sup>4</sup> TCID <sub>50</sub> /mL	No (3/3 negative)	No (3/3 positive)
Parainfluenza virus 2	1.6 x 10 <sup>5</sup> TCID <sub>50</sub> /mL	No (3/3 negative)	No (3/3 positive)
Parainfluenza virus 3	1.6 x 10 <sup>6</sup> TCID <sub>50</sub> /mL	No (3/3 negative)	No (3/3 positive)
Parainfluenza virus 4	1.15×10 <sup>5</sup> TCID <sub>50</sub> /mL	No (3/3 negative)	No (3/3 positive)
Influenza A NP antigen	180 µg/mL	No (3/3 negative)	No (3/3 positive)
Influenza B NP antigen	200 µg/mL	No (3/3 negative)	No (3/3 positive)
Enterovirus	2.8 x 10 <sup>5</sup> TCID <sub>50</sub> /mL	No (3/3 negative)	No (3/3 positive)

Respiratory syncytial virus	2.8 x 10 <sup>4</sup> TCID <sub>50</sub> /mL	No (3/3 negative)	No (3/3 positive)
Rhinovirus	2.2 x 10 <sup>5</sup> PFU/mL	No (3/3 negative)	No (3/3 positive)
Haemophilus influenzae	5.2 x 10 <sup>5</sup> CFU/mL	No (3/3 negative)	No (3/3 positive)
Streptococcus pneumoniae	>2×10 <sup>3</sup> CFU/mL	No (3/3 negative)	No (3/3 positive)
Streptococcus pyogenes	3.6 x 10 <sup>5</sup> CFU/mL	No (3/3 negative)	No (3/3 positive)
Candida albicans	4.5×10 <sup>6</sup> TCID <sub>50</sub> /mL	No (3/3 negative)	No (3/3 positive)
Pooled human nasal wash – representative of normal respiratory microbial flora	N/A	No (3/3 negative)	No (3/3 positive)
Bordetella pertussis	3.9 x 10 <sup>7</sup> CFU/mL	No (3/3 negative)	No (3/3 positive)
Mycoplasma pneumoniae	4.4 x 10 <sup>5</sup> CFU/mL	No (3/3 negative)	No (3/3 positive)
Chlamydia pneumoniae	1.4 x 10 <sup>7</sup> IFU/mL	No (3/3 negative)	No (3/3 positive)
Legionella pneumophila	7.8 x 10 <sup>5</sup> CFU/mL	No (3/3 negative)	No (3/3 positive)
Mycobacterium tuberculosis	>2×10 <sup>3</sup> CFU/mL	No (3/3 negative)	No (3/3 positive)
Pneumocystis jirovecii (PJP)	3.45×10 <sup>6</sup> CFU/mL	No (3/3 negative)	No (3/3 positive)

**3. Interfering Substances**

The following potentially interfering substances, naturally present in respiratory specimens or that may be artificially introduced into the nasal cavity or nasopharynx, were evaluated with the OnSite COVID-19 Ag Rapid Test at the concentrations listed in the following table and were found not to affect test performance for detection of both positive and negative specimens:

Interfering Substance	Concentration	Interference (Yes/No)	Interfering Substance	Concentration	Interference (Yes/No)
Mucin	0.5%	No (6/6 correct)	Ribavirin	1 mg/mL	No (6/6 correct)
Whole Blood	4%	No (6/6 correct)	Peramivir	1 mg/ml	No (6/6 correct)
Phenylephrine	15% v/v	No (6/6 correct)	Tobramycin	4 µg/mL	No (6/6 correct)
Fluconazole	5% w/v	No (6/6 correct)	Diphenhydramine	0.08 mg/dL	No (6/6 correct)
Budesonide	5% w/v	No (6/6 correct)	Dextromethorphan	1.56 µg/dL	No (6/6 correct)
Nasal Gel	2% v/v	No (6/6 correct)	Acetaminophen	199 µM	No (6/6 correct)
Menthol	1.5 mg/mL	No (6/6 correct)	Acetylsalicylic Acid	3 mg/dL	No (6/6 correct)
Benzocaine	1.5 mg/mL	No (6/6 correct)	Mupirocin	10 mg/mL	No (6/6 correct)
Lopinavir	5 mg/mL	No (6/6 correct)	HAMA	4 ng/mL	No (6/6 correct)
Zanamivir	5 mg/mL	No (6/6 correct)	Biotin	100 µg/mL	No (6/6 correct)
Oseltamivir	5 mg/mL	No (6/6 correct)			

**4. Hook Effect**

No high dose hook effect was observed when tested with up to a concentration of 3×10<sup>8</sup> pg/mL of recombinant SARS-CoV-2 NP antigen with the OnSite COVID-19 Ag Rapid Test.

**LIMITATIONS OF TEST**

- The Assay Procedure and the Interpretation of Assay Result must be followed closely when testing for the presence of SARS-CoV-2 antigen in the swab specimen from individual subjects. For optimal test performance, proper sample collection is critical. Failure to follow the procedure may lead to inaccurate results.
- It is intended for use only by healthcare professionals or personnel trained in rapid test procedure. For *in vitro* diagnostic use only.
- The OnSite COVID-19 Ag Rapid Test is limited to the qualitative detection of SARS-CoV-2 antigen. The intensity of the test line does not have linear correlation with virus titer in the specimen.
- Sensitivity can differ with various strains of SARS-CoV-2 due to differences of antigen expression. Specimens might contain a new or non-identified strain of SARS-CoV-2 that expresses varying amounts of antigen.
- A negative or non-reactive result for an individual subject indicates absence of detectable SARS-CoV-2 antigen. However, a negative or non-reactive result does not preclude the possibility of SARS-CoV-2 virus infection.
- A negative or non-reactive result can occur if the quantity of the SARS-CoV-2 virus (antigen) present in the specimen is below the detection limit of the assay, or if the virus detected was not present in the swab specimen sampled, or the viruses have undergone minor amino acid mutation in the epitope recognized by the antibody utilized in the test.
- The OnSite COVID-19 Ag Rapid Test detects both viable and non-viable SARS-CoV and SARS-CoV-2 antigens. Test performance depends on antigen loaded in the sample. A positive test does not rule out the possibility that other pathogens may be present.
- Performance of the OnSite COVID-19 Ag Rapid Test has been validated in specimens stored in multiple viral transport media (VTM). However, specimens stored in PBS or saline solutions should not be tested on the OnSite COVID-19 Ag Rapid Test.
- Performance of the test has not been established for monitoring antiviral treatment of SARS-CoV-2 infection.

**REFERENCES**

- Naming the coronavirus disease (COVID-19) and the virus that causes it. (n.d.). Retrieved from [https://www.who.int/emergencies/diseases/novel-coronavirus-2019/technical-guidance/naming-the-coronavirus-disease-\(covid-2019\)-and-the-virus-that-causes-it](https://www.who.int/emergencies/diseases/novel-coronavirus-2019/technical-guidance/naming-the-coronavirus-disease-(covid-2019)-and-the-virus-that-causes-it).
- "Frequently Asked Questions - General Assembly of the United Nations." United Nations, [www.un.org/pga/75/coronavirus/faqs/](http://www.un.org/pga/75/coronavirus/faqs/).
- World Health Organization. (2020). Advice on the use of masks in the community, during home care, and in health care settings in the context of COVID-19: interim guidance, 19 March 2020 (No. WHO/2019-nCoV/IPC\_Masks/2020.2). World Health Organization.
- Healthcare Professionals: Frequently Asked Questions and Answers. (2020, March 22). Retrieved from <https://www.cdc.gov/coronavirus/2019-ncov/hcp/faq.html>.

**Index of Symbols**

	Consult instructions for use		For <i>in vitro</i> diagnostic use only		Use by
	Catalog #		Lot Number		Tests per kit
	Store between 2-30°C		Authorized Representative		Do not reuse
	Manufacturer		Date of manufacture		

**CTK Biotech, Inc.**  
 13855 Stowe Drive,  
 Poway, CA 92064, USA  
 Tel: 858-457-8698  
 Fax: 858-535-1739  
 E-mail: [info@ctkbiotech.com](mailto:info@ctkbiotech.com)

**MDSS GmbH**  
 Schiffgraben 41  
 30175 Hannover, Germany

PI-R0182C-FS Rev. D  
 Date released: 2021-05-13  
 English version

For Export Only, Not For Re-sale in the USA.