

Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Temocillin TMO 30 µg	111721063	2024.11.16

Ref.

9186 – 9186/1 – 9186/2

Antimicrobial Susceptibility Testing

Tested according to current EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	16-22	19

Batch Release

Approved

Date

23.11.2021

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.