

Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Cefotaxime+Clavulanic acid CTL 40 (30+10) µg	011422083	2025.01.13

Ref.	Initial
9182 – 9182/1 – 9182/2	CTL

Control strains	Medium	Inoculum	Incubation	Results Cefotaxime 30µg zone (mm)	Results Cefotaxime+ Clavulanic acid 40 µg zone (mm)
<i>Escherichia coli</i> ATCC 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C ambient 16-18 h	31	31
<i>Escherichia coli</i> DSMZ 22664	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C ambient 16-18 h	15	30
<i>Escherichia coli</i> DMSZ 22665	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C ambient 16-18 h	11	26
<i>Klebsiella pneumoniae</i> ATCC 700603	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C ambient 16-18 h	18	27

Batch Release

Approved

Date

21.01.2022

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.